

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply for Building-In

Name and address of the applicant

 Delta Electronics (Thailand) Public Co., Ltd.
 909 Soi 9, Moo 4, Bangpoo Industrial Estate (E.P.Z.),
 Pattana 1 Road, Tambol Phraksa, Amphur Muang,
 Samutprakarn 10280, Thailand

Name and address of the manufacturer

 Delta Electronics (Thailand) Public Co., Ltd.
 909 Soi 9, Moo 4, Bangpoo Industrial Estate (E.P.Z.),
 Pattana 1 Road, Tambol Phraksa, Amphur Muang,
 Samutprakarn 10280, Thailand

Name and address of the factory

See next pages

Note: When more than one factory, please report on page 2
 Additional Information on page 2

Ratings and principal characteristics

See next pages

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF-2

Model / Type Ref.

See next pages

Additional information (if necessary may also be reported on page 2)

 Unit also complies with EN 62368-1:2014 + A11:2017
 and BS EN 62368-1:2014 + A11:2017

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 62368-1:2014

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0612/21 (2021-10-12)

This CB Test Certificate is issued by the National Certification Body


 SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia
 T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

Date: 2021-10-12

Signature: Bojan Pečavar



Name and address of the factory:

Delta Electronics (Thailand) Public Co., Ltd.
909 Soi 9, Moo 4, Bangpoo Industrial Estate (E. P. Z),
Pattana 1 Rd., Tambol Phraksa, Amphur Muang,
Samutprakarn 10280, Thailand, India

Model names, ratings and principal characteristics:

Model names:

IMA-S2000-24-YVxxx (x=0,9, A-Z or blank),
IMA-C2000-24-YVxxx (x=0,9, A-Z or blank)

The symbols "xxx" in the model names can be (x=0-9, A-Z or blank)

Input:

100-127 Vac; 17 A – 15 A; 50/60 Hz
200-240 Vac; 12 A - 9.8 A; 50/60 Hz
120-300 Vdc; 12 A - 7.8 A

Output:

DC 24 V; 60,41 A Max @ 100-127 Vac
DC 24 V; 83,4 A Max @ 200-240 Vac
DC 24 V; 41,7 A Max @ 120-240 Vdc
DC 24 V; 83,4 A Max @ 280-300 Vdc
DC 5 V; 2 A Max
Total power 2010 W Max.

Additional information (if necessary)

Date: 2021-10-12

Signature: Bojan Pečavar



IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply for Building-In

Name and address of the applicant

Delta Electronics (Thailand) Public Co., Ltd.
909 Soi 9, Moo 4, Bangpoo Industrial Estate (E.P.Z.),
Pattana 1 Road, Tambol Phraksa, Amphur Muang,
Samutprakarn 10280, Thailand

Name and address of the manufacturer

Delta Electronics (Thailand) Public Co., Ltd.
909 Soi 9, Moo 4, Bangpoo Industrial Estate (E.P.Z.),
Pattana 1 Road, Tambol Phraksa, Amphur Muang,
Samutprakarn 10280, Thailand

Name and address of the factory

Delta Electronics (Thailand) Public Co., Ltd.
909 Soi 9, Moo 4, Bangpoo Industrial Estate (E. P. Z),
Pattana 1 Rd., Tambol Phraksa, Amphur Muang,

Note: When more than one factory, please report on page 2

 Additional Information on page 2

Ratings and principal characteristics

See next page.

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

CTF-2

Model / Type Ref.

IMA-S2000-24-YVxxx (x=0,9, A-Z or blank),
IMA-C2000-24-YVxxx (x=0,9, A-Z or blank)

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 + A1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0596/21 (2021-10-20)

This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

Date: 2021-10-20

Signature: Zoran Svetik 

Ratings and principal characteristics:

Input:

100-127 Vac; 17 A-15 A; 50/60 Hz

200-240 Vac; 12 A-9,8 A; 50/60 Hz

Output:

24 Vdc; 60,41 A Max @ 100-127 Vac

24 Vdc; 83,4 A Max @ 200-240 Vac

5 Vdc; 2 A Max

Total power 2010 W Max.

Additional information (if necessary)

Date: 2021-10-20

Signature: Zoran Svetik 