



# Certificate of Compliance

**Certificate:** 80051455

**Master Contract:** 158114 (LR 10099)

**Project:** 80051455

**Date Issued:** 2020-08-31

**Issued to:** Delta Electronics, Inc.  
3 Tung Yuan Rd,  
Chungli Industrial Zone,  
Taoyuan City, 3063  
TAIWAN

**Attention:** Mr. Elvis Kuo

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only*



*Alfred Lee*  
**Issued by:** Alfred Lee

## PRODUCTS

CLASS 5311-28 - POWER SUPPLIES - Component Type - For Use in Medical Equipment/System (Adopted IEC 60601-1 3rd edition)

CLASS 5311-98 - POWER SUPPLIES - Component Type - For Use in Medical Equipment/System - Certified to US Standards (Adopted IEC 60601-1 3rd edition)

Switching Power Supply (AC to DC Configurable Power Supply), Model: MEG-700A3

### Rated Input:

100-240V~, 8.5A-3.5A, 50-60Hz, Connected to the AC mains supply by Appliance inlet or Primary terminal block.

### Rated Output:

700W MAX. (SLOT1-SLOT3); DC 5Vsb/2A

See below table for output rated voltage and rated current when different type of output module board be inserted into each slot of power supply.



**Certificate:** 80051455  
**Project:** 80051455

**Master Contract:** 158114  
**Date Issued:** 2020-08-31

**Single Slot, Single Output module:**

Code	Output Module board name	Output V (d.c.)	Output current (A)	Output watts (W)
A1	MEG-A090M02S AA	2	45	90
B1	MEG-A108MS02 AA	2.4	45	108
C1	MEG-A135M03S AA	3	45	135
D1	MEG-A149M03S AA	3.3	45	149
E1	MEG-A225M05S AA	5	45	225
F1	MEG-A248M06S AA	5.5	45	248
G1	MEG-A252M06S AA	6	42	252
H1	MEG-A200M08S AA	8	25	200
I1	MEG-A250M10S AA	10	25	250
J1	MEG-A300M12S AA	12	25	300
K1	MEG-A300M14S AA	14	21.4	300
L1	MEG-A300M15S AA	15	20	300
M1	MEG-A300M18S AA	18	16.7	300
N1	MEG-A300M20S AA	20	15	300
O1	MEG-A300M24S AA	24	12.5	300
P1	MEG-A300M28S AA	28	10.7	300
Q1	MEG-A300M30S AA	30	10	300
R1	MEG-A300M32S AA	32	9.4	300
S1	MEG-A300M36S AA	36	8.3	300
T1	MEG-A300M42S AA	42	7.1	300
U1	MEG-A300M48S AA	48	6.3	300
V1	MEG-A300M54S AA	54	5.5	300
W1	MEG-A300M60S AA	60	5	300

**Single Slot, Dual Outputs module:**

Code	Output Module board name	Output V (d.c.)	Output current (A)	Output watts (W)
D	MEG-A033M03D AA	3.3	5	16.5
E	MEG-A050M05D AA	5	5	25
F	MEG-A055M05D AA	5.5	5	27.5
G	MEG-A060M06D AA	6	5	30
H	MEG-A080M08D AA	8	5	40
I	MEG-A100M10D AA	10	5	50
J	MEG-A120M12D AA	12	5	60
K	MEG-A140M14D AA	14	5	70
L	MEG-A150M15D AA	15	5	75
M	MEG-A180M18D AA	18	5	90
N	MEG-A200M20D AA	20	4.5	90
O	MEG-A192M24D AA	24	3.7	90
P	MEG-A224M28D AA	28	3.2	90
Q	MEG-A240M30D AA	30	3A	90

**Certificate:** 80051455  
**Project:** 80051455

**Master Contract:** 158114  
**Date Issued:** 2020-08-31

**Triple Slot, Single Output module:**

Code	Output Module board name	Output V (d.c.)	Output current (A)	Output watts (W)
H2	MEG-A800M08S AA	8	78.75	630
I2	MEG-A1K0M10S AA	10	70	700
J2	MEG-A1K2M12S AA	12	58.33	700
K2	MEG-A1K2M14S AA	14	50	700
L2	MEG-A1K2M15S AA	15	46.67	700
M2	MEG-A1K2M18S AA	18	38.89	700
N2	MEG-A1K2M20S AA	20	35	700
O2	MEG-A1K2M24S AA	24	29.17	700
P2	MEG-A1K2M28S AA	28	25	700
Q2	MEG-A1K2M30S AA	30	23.33	700
R2	MEG-A1K2M32S AA	32	21.88	700
S2	MEG-A1K2M36S AA	36	19.45	700
T2	MEG-A1K2M42S AA	42	16.67	700
U2	MEG-A1K2M48S AA	48	14.59	700
V2	MEG-A1K2M54S AA	54	12.97	700
W2	MEG-A1K2M60S AA	60	11.67	700

Up to three sets connector (Slot 1 to Slot 3) provided in main board for installation various kind of output module board to provide different output voltage and current level as descriptive in above table.

1. Medical device protection against electric shock: Class I
2. Applied Part protection against electric shock: No applied part
3. Degree of protection against ingress of water or particulate matter: IPX0
4. Method of Sterilization: None
5. Suitability for use in an Oxygen Rich Environment: Medical device not intended to be used in an Oxygen Rich Environment
6. Suitability to use Medical device in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide: Medical device not suitable for use in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide.
7. Mode of operation: Continuous
8. Environmental Conditions: Normal: -20°C to 50°C, 5 to 95% RH, 54-106kPa as specified by manufacturer.

**APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 60601-1:14 Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance

ANSI/AAMI ES60601-1:2005/(R) 2012, AND C1:2009 AND A2:2010(R) 2012 Medical electrical equipment - Part 1: General requirements for basic safety and essential performance



## *Supplement to Certificate of Compliance*

**Certificate:** 80051455

**Master Contract:** 158114 (LR 10099)

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80051455	2020-08-31	Original Certification.